



Certificate of Analysis

Product Compact Dry X-BC 240plates (Code 06728)

Lot No 122002

Expiry Date 2021 . 7

	Specification	Results
Appearance	Light yellow . No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 6 %
Aseptic Test	No growth of any bacteria.	Pass
pH	6.8 ~ 7.2	7 . 0

Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 33-37°C for 22-26hours, good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Bacillus cereus</i>	ATCC 19637	+	blue-green
<i>Bacillus cereus</i>	ATCC 11778	+	blue-green

- ② When following test strains are inoculated to incubate at 33-37°C for 22-26hours, growths should be inhibited.

<u>Test strain</u>		<u>Growth</u>
<i>Enterococcus faecalis</i>	ATCC 19433	-
<i>Bacillus subtilis</i>	ATCC 6633	-
<i>Staphylococcus aureus</i>	ATCC6538	-
<i>Escherichia coli</i>	ATCC 8739	-
<i>Pseudomonas aeruginosa</i>	ATCC9027	-

Judge: Acceptable quality.

Certificate issue date : 2020 . 2 . 17

Signed (QC Manager)