



Certificate of Analysis

Product Compact Dry YM 1400plates ( Code 54054 )
Lot No 461810
Expiry Date 2020 . 3

Table with 3 columns: Specification, Results. Rows include Appearance (Pass), Loss of drying (4.0%), Aseptic Test (Pass), pH (5.2 ~ 5.8).

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 25C for 3-7days, good growths and color developments should be observed.

Table with 3 columns: Test strain, Growth, Color development. Rows include Candida albicans ATCC2091, ATCC10231 and Aspergillus brasiliensis ATCC16404.

When following test strains are inoculated to incubate at 25C for 3-7days, growths should be inhibited.

Table with 2 columns: Test strain. Rows include Bacillus subtilis ATCC6633 and Escherichia coli ATCC8739.

Judge: Acceptable quality.

Certificate issue date : 2018 . 10 . 24

Signed (Chief Operator) Emiko Takahashi

Signed (QA Manager) Tadanobu Kurozuka