



Certificate of Analysis

Product Compact Dry YM 1400plates (Code 54054)

Lot No 467903

Expiry Date 2020 . 8

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 8 %
Aseptic Test	No growth of any bacteria.	Pass
pH	5.2 ~ 5.8	5 . 6

Evaluation of Performance Specification

①When following test strains are inoculated to incubate at 25°C for 3-7days,good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Candida albicans</i> ATCC2091	+	light bluish green
<i>Candida albicans</i> ATCC10231	+	light bluish green
<i>Aspergillus brasiliensis</i> ATCC16404	+	blue

②When following test strains are inoculated to incubate at 25°C for 3-7days, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Escherichia coli</i> ATCC8739

Judge: Acceptable quality.

Certificate issue date :2019 . 3 . 26

Signed (QC Manager)