



Certificate of Analysis

Product Compact Dry CF 1400plates (Code 54053)

Lot No 206810

Expiry Date 2020 . 3

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 6 %
Aseptic Test	No growth of any bacteria.	Pass
pH	6.8 ~ 7.2	7 . 0

Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 35°C for 20-24hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC8739	bluish green
<i>Klebsiella oxytoca</i> ATCC13182	bluish green
<i>Pseudomonas aeruginosa</i> ATCC9027	white

- ② When following test strains are inoculated to incubate at 35°C for 20-24hours, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Staphylococcus aureus</i> ATCC6538

Judge: Acceptable quality.

Certificate issue date : 2018 . 10 . 19

Signed (Chief Operator)

Signed (QA Manager)