



## Certificate of Analysis

Product      Compact Dry CF 1400plates    ( Code 54053 )  
 Lot No        210001  
 Expiry Date   2021 . 12

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	5 . 4 %
Aseptic Test	No growth of any bacteria.	Pass
pH	6.8 ~ 7.2	7 . 0

Evaluation of Performance  
 Specification

① When following test strains are inoculated to incubate at 35°C for 20-24hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC8739	bluish green
<i>Klebsiella oxytoca</i> ATCC13182	bluish green
<i>Pseudomonas aeruginosa</i> ATCC9027	white

② When following test strains are inoculated to incubate at 35°C for 20-24hours, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Staphylococcus aureus</i> ATCC6538

Judge:    Acceptable quality.

Certificate issue date : 2020 . 2 . 3

Signed (QC Manager)

*Takafumi Kato*