



# Certificate of Analysis

Product      Compact Dry EC 1400plates    ( Code 54052 )  
 Lot No        450901  
 Expiry Date   2020 . 6

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 7 %
Aseptic Test	No growth of any bacteria.	Pass
pH	6.8 ~ 7.2	6 . 9

Evaluation of Performance  
 Specification

① When following test strains are inoculated to incubate at 35°C for 20-24hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC8739	+	bluish green
<i>Klebsiella oxytoca</i> ATCC13182	+	reddish purple
<i>Pseudomonas aeruginosa</i> ATCC9027	+	white

② When following test strains are inoculated to incubate at 35°C for 20-24hours, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Staphylococcus aureus</i> ATCC6538

Judge:    Acceptable quality.

Certificate issue date :   2019 . 1 . 30

Signed (Chief Operator)    *Emiko Takahashi*

Signed (QA Manager)        *Tadanobu Kirschuka*