



Certificate of Analysis

Product Compact Dry LS 1400plates (Code 54060)

Lot No 222811

Expiry Date 2020 . 4

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 6 %
Aseptic Test	No growth of any bacteria.	Pass
pH	7.0 ~ 7.4	7 . 4

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 35°C for 22-26hours, good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Listeria monocytogenes</i>	JCM 7680	+	blue-green
<i>Listeria monocytogenes</i>	Lm-1	+	blue-green
<i>Enterococcus faecalis</i>	ATCC 19433	-	
<i>Bacillus subtilis</i>	ATCC 6633	-	
<i>Escherichia coli</i>	ATCC 8739	-	
<i>Pseudomonas aeruginosa</i>	ATCC 9027	-	

Judge: Acceptable quality.

Certificate issue date : 2018 . 11 . 8

Signed (Chief Operator)

Emiko Takahashi

Signed (QA Manager)

Tadanobu Kurozuka