



Certificate of Analysis

Product Compact Dry TC 1400plates (Code 54051)

Lot No 842812

Expiry Date 2020 . 5

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	5 . 4 %
Aseptic Test	No growth of any bacteria.	Pass
pH	6.8 ~ 7.2	7 . 0

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 35°C for 44-48hours,good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Escherichia coli</i>	ATCC 8739	+	red
<i>Klebsiella oxytoca</i>	ATCC 13182	+	red
<i>Pseudomonas aeruginosa</i>	ATCC 9027	+	red
<i>Staphylococcus aureus</i>	ATCC 6538	+	red
<i>Bacillus subtilis</i>	ATCC 6633	+	red
<i>Lactobacillus lactis</i>	JCM 6123	+	red

Judge: Acceptable quality.

Certificate issue date : 2018 . 12 . 19

Signed (Chief Operator)

Emiko Takahashi

Signed (QA Manager)

Tadanobu Kirsyuka