



Certificate of Analysis

Product Compact Dry VP 240plates (Code 06749)

Lot No 115808

Expiry Date 2020 . 1

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 8 %
Aseptic Test	No growth of any bacteria.	Pass
pH	8.6 ~ 9.0	9 . 0

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 35 for 20-24hours, good growths and color should be observed as follows;

<u>Test strain</u>	<u>Growth</u>	<u>Color</u>
<i>Vibrio vulnificus</i> ATCC29307	+	Magenta
<i>Vibrio parahaemolyticus</i> ATCC17082	+	Blue/Blue-green
<i>Vibrio alginolyticus</i> 18	+	Milk white
<i>Vibrio cholerae</i> 3898	+	Magenta
<i>Klebsiella oxytoca</i> ATCC13182	-	-
<i>Pseudomonas aeruginosa</i> ATCC27853	-	-
<i>Enterococcus faecalis</i> ATCC19433	-	-

Judge: Acceptable quality.

Certificate issue date : 2018 . 8 . 29

Signed (Chief Operator)

Emiko Takahashi

Signed (QA Manager)

Tadanobu Hirozuka