



Certificate of Analysis

Product Compact Dry VP 240plates (Code 06749)

Lot No 125009

Expiry Date 2022 . 2

| | Specification | Results |
|----------------|---|---------|
| Appearance | Light yellow. No particular matters. | Pass |
| Loss of drying | Less than 10.0% | 4 . 4 % |
| Aseptic Test | No growth of any bacteria. | Pass |
| pH | 8.6 ~ 9.0 | 9 . 0 |

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 35°C for 18 - 20 hours,
good growths and color should be observed as follows;

| <u>Test strain</u> | <u>Growth</u> | <u>Color</u> |
|--|---------------|-----------------|
| <i>Vibrio vulnificus</i> ATCC29307 | + | Magenta |
| <i>Vibrio parahaemolyticus</i> ATCC17082 | + | Blue/Blue-green |
| <i>Vibrio alginolyticus</i> 18 | + | Milk white |
| <i>Vibrio cholerae</i> 3898 | + | Magenta |
| <i>Klebsiella oxytoca</i> ATCC13182 | - | - |
| <i>Pseudomonas aeruginosa</i> ATCC27853 | - | - |
| <i>Enterococcus faecalis</i> ATCC19433 | - | - |

Judge: Acceptable quality.

Certificate issue date :2020 . 10 . 12

Signed (QC Manager)