



# Certificate of Analysis

Product      Compact Dry X-SA 1400plates    ( Code 54057 )  
 Lot No        321812  
 Expiry Date   2020 . 5

	Specification	Results
Appearance	Straw color. No particular matters.	Pass
Loss of drying	Less than 10.0%	5 . 4 %
Aseptic Test	No growth of any bacteria.	Pass
pH	7.2 ~ 7.6	7 . 5

Evaluation of Performance  
 Specification

① When following test strains are inoculated to incubate at 36°C for 22-26hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Staphylococcus aureus</i> ATCC6538	+	blue
<i>Staphylococcus aureus</i> ATCC12600	+	blue
<i>Staphylococcus epidermidis</i> ATCC14990	+	white

② When following test strains are inoculated to incubate at 36°C for 22-26hours, growths should be inhibited.

<u>Test strain</u>
<i>Enterococcus faecalis</i> ATCC19433
<i>Escherichia coli</i> ATCC8739
<i>Pseudomonas aeruginosa</i> ATCC9027

Judge:    Acceptable quality.

Certificate issue date :   2019 . 1 . 11

Signed (Chief Operator)    *Emiko Takahashi*

Signed (QA Manager)        *Tadanobu Kishizuka*