



# Certificate of Analysis

Product Compact Dry X-SA 1400plates ( Code 54057 )

Lot No 325902

Expiry Date 2020 . 7

	Specification	Results
Appearance	Straw color. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 6 %
Aseptic Test	No growth of any bacteria.	Pass
pH	7.2 ~ 7.6	7 . 5

## Evaluation of Performance Specification

① When following test strains are inoculated to incubate at 36°C for 22-26hours,good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Staphylococcus aureus</i> ATCC6538	+	blue
<i>Staphylococcus aureus</i> ATCC12600	+	blue
<i>Staphylococcus epidermidis</i> ATCC14990	+	white

② When following test strains are inoculated to incubate at 36°C for 22-26hours, growths should be inhibited.

<u>Test strain</u>
<i>Enterococcus faecalis</i> ATCC19433
<i>Escherichia coli</i> ATCC8739
<i>Pseudomonas aeruginosa</i> ATCC9027

Judge: Acceptable quality.

Certificate issue date : 2019 . 3 . 14

Signed (QC Manager)