



Certificate of Analysis

Product Compact Dry YMR 1400plates (Code 54063)

Lot No 020807

Expiry Date 2019 . 12

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 1 %
Aseptic Test	No growth of any bacteria.	Pass
pH	5.2 ~ 5.8	5 . 6

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 25 for 2-3days, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Candida albicans</i> ATCC10231	+	light yellow green
<i>Candida glabrata</i> JCM3699	+	bluish green
<i>Aspergillus brasiliensis</i> ATCC16404	+	blue

When following test strains are inoculated to incubate at 25 for 2-3days, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Escherichia coli</i> ATCC8739

Judge: Acceptable quality.

Certificate issue date : 2018 . 7 . 27

Signed (Chief Operator)

Signed (QA Manager)